


OrthoCAPE

Surgical technique Varisation staples

(The specific ancillary equipment is required for fitting ORTHO CAPE implants)



- 1 Make one distal hole in the distal fragment, directed forward and outward
- 2 Insert the staple with the straight staple driver (90°) or the oblique staple driver (26°)
- 3 Finish with the impactor (90° or 26°)

